

Set de verificación para Checker HC pH línea piscinas HI779-11



Certificate of Analysis

Product name: CERTIFIED STANDARD CUVETTES
 pH
 Product code: HI779-11
 Lot number: SC0991/20
 Best used before: December 2023
 Issuing date: 2020.12.08

Standard Cuvettes Specifications @ 25°C:

Standard cuvette code	Lot number	Value	Analysis Results
HI779-11A	594		Passed
HI779-11B	594		Passed

Hanna Instruments certifies that the above information is true and correct. These should be used for general calibration purposes only. Do not open standard cuvettes until ready for use.
 Instructions for use:
 1. Thoroughly wipe surface of cuvette. The cuvette should be used immediately after opening top of cuvette. The cuvette should be used immediately after opening top of cuvette.
 2. At the C.I. press operation.
 3. At the C.I. press operation.
 Accepted by: _____
 Title: _____
 Signature: _____
 Date: _____

CoA_HI7xx-11_3_rev.0.1_November 2020

Hanna Instruments, Inc., 584 Park East Drive, Woonsocket, RI 02895, USA
 www.hannainst.com

Descripción

El Set de verificación de la línea piscinas Checker **HI779-11** entrega una solución simple para la evaluación de desempeño del colorímetro de bolsillo **HI779**. Este set de estándares de alta calidad se desarrolló en nuestra fábrica especializada y cuenta con certificado de análisis. El certificado de análisis brinda información del número de lote, valores de referencia y fechas de vencimiento para la trazabilidad en las verificaciones del Checker HC.

Especificaciones

Certificado de análisis	Si
Empaque	Cubetas
Cantidad	1 set

Accesorios

No Especifica

Cómo pedir

El **HI 779-11** se entrega junto a las cubetas certificadas de pH A - **HI779-11A** (1), cubetas certificadas de pH B - **HI 779-11B** (1), certificado de análisis, y manual de instrucciones.

Ventajas

Los estándares **HI779-11** son un set de alta calidad que permite a los usuarios validar el desempeño de los checkers de pH de la línea de piscinas **HI779**. El **HI779-11** incluye una cubeta cero sellada (muestra sin reaccionar conocida como blanco) y una cubeta teñida con el valor estándar.

-Estándares prefabricados de pH con valor 8.0 pH +/- 0.1 pH @ 25°C.

-Reutilizables y con amplia vida útil.

-Se entrega junto al certificado de análisis:

- Código de producto
- Número de lote
- Fecha de vencimiento



CERTIFICATE OF ANALYSIS

Product Name: CALIBRATION CHECKING SET FOR HI 711
 Product Code: HI 711-11
 Lot number: SC0372/14
 Expiration date: October 2017

Standard cuvette	Lot number	Standard Value @ 25°C (ppm)	Passed
A, HI 711-11	2038	1.00 ± 0.05	✓
B, HI 711-11	2039	1.00 ± 0.05	✓

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not use standard cuvettes.


Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
2. At the C.1 prompt place A, HI 711-11 standard cuvette into the holder of the Checker and press operations button.
3. At the C.2 prompt place B, HI 711-11 standard cuvette into the holder and press the operations button.

Acceptable results must be between 0.95 and 1.05 ppm @ 25°C.

File number: CERT711-11_SC0372/14

QA manager: Bogdan Munteanu



Video

No Especifica