

**Set de Verificación para
 Checker HC Cobre Rango
 Alto Línea de Piscinas**
 HI7024-11



Certificate of Analysis

Product name: CERTIFIED STANDARD CUVETTES
 Copper High Range

Product code: HI7024-11

Lot number: SC0990/20

Best used before: June 2023

Issuing date: 2020.12.08

Standard Cuvettes Specifications @ 25°C:

Standard cuvette code	Lot number	Analysis Results
HI7024-11A	5947	Passed
HI7024-11B	5947	Passed

Hanna Instruments certifies the...
 These should be used for quality control...
 Hanna instrument only. Store...
 Do not open standard cuvette...
 Instructions for use:
 1. Thoroughly wipe standard cuvette...
 2. At the C.1 prompt...
 3. At the C.2 prompt...
 operations...
 Acceptable...
 Quality Control Manager / Engineer
 (Signature)
 CoA_HI7xx-11_3_rev.0.1_November 2020
 CERTPL105_11/20

Hanna Instruments, Inc., 584 Park East Drive, Woonsocket, RI 02895, USA
 www.hannainst.com

Descripción

El set de verificación de la línea de piscinas Checker **HI 7024-11** entrega una solución simple para la evaluación de desempeño del colorímetro de bolsillo **HI 7024** para cobre rango alto en un rango 0.00 a 5.00 ppm. Este set de estándares de alta calidad se desarrolló en nuestra fábrica especializada y cuenta con certificado de análisis. El certificado de análisis brinda información del número de lote, valores de referencia y fechas de vencimiento para la trazabilidad en las verificaciones del Checker HC.

Especificaciones

Certificado de análisis	Si
Empaque	Cubetas
Cantidad	1 set

Accesorios

No Especifica

Cómo pedir

El **HI 7024-11** se entrega junto a las cubetas certificadas de cobre A - **HI7014-11A** (1), cubetas certificadas de cobre B - **HI7014-11B** (1), certificado de análisis, y manual de instrucciones.

Ventajas

Los estándares **HI 7024-11** son un set de alta calidad que permite a los usuarios validar el desempeño de los checkers de cobre rango alto de la línea de piscinas. El **HI7024-11** incluye una cubeta cero sellada y una cubeta teñida con valor 2.00 +/- 0.10 ppm, cobre @25°C. El set de verificación se produce en nuestra fábrica especializada utilizando reactivos de grado investigativo con ambientes a temperatura controlada.

El HI7024-11 se entrega junto al certificado de análisis

- Código de producto
- Número de lote
- Fecha de vencimiento



CERTIFICATE OF ANALYSIS

Product Name: CALIBRATION CHECKING SET FOR HI 711
 Product Code: HI 711-11
 Lot number: SC0372/14
 Expiration date: October 2017

Standard cuvette	Lot number	Standard Value @ 25°C (ppm)	Passed
A, HI 711-11	2038	1.00 ± 0.05	✓
B, HI 711-11	2039	1.00 ± 0.05	✓

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not use standard cuvettes.


Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
2. At the C.1 prompt place A, HI 711-11 standard cuvette into the holder of the Checker and press operations button.
3. At the C.2 prompt place B, HI 711-11 standard cuvette into the holder and press the operations button.

Acceptable results must be between 0.95 and 1.05 ppm @ 25°C.

File number: CERT711-11_SC0372/14

QA manager: Bogdan Munteanu



Video

No Especifica